

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.		
10/696,664 10/29/2003		Rosalinda C. Sta-Maria	Sta-Maria-I	6832		
7590 01/18/2006		EXAM	EXAMINER			
Mr. Walter J. Tencza Jr.			PATEL, N	PATEL, MITAL B		
Suite 3						
10 Station Place	•	ART UNIT	PAPER NUMBER			
Metuchen, NJ 08840			3743			

DATE MAILED: 01/18/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary		Application	No.	Applicant(s) STA-MARIA, ROSALINDA C.				
		10/696,664						
		Examiner		Art Unit				
		Mital B. Pat		3743				
 Period for	The MAILING DATE of this communic Reply	cation appears on the c	over sheet with the	correspondence ad	dress			
WHICH - Extens after S - If NO p - Failure Any re	PRTENED STATUTORY PERIOD FOR HEVER IS LONGER, FROM THE MAIN HE MAIN HE WAS A STATE OF THE WAS A STATE OF THE MAIN HE WAS A STATE OF THE MAIN HE WAS A STATE OF THE MAIN HE WAS A STATE OF THE WAS A STATE OF THE MAIN HE WAS A STATE OF THE WAS A STATE OF	AILING DATE OF THIS of 37 CFR 1.136(a). In no eventualization. tutory period will apply and will ovill, by statute, cause the application.	S COMMUNICATIO I, however, may a reply be tile expire SIX (6) MONTHS from ation to become ABANDONE	N. mely filed n the mailing date of this co ED (35 U.S.C. § 133).				
Status								
1) X F	Responsive to communication(s) filed	d on 10 November 200)5.					
•		b) This action is not						
•	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is							
• —	closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.							
Dispositio	n of Claims							
4)⊠ Claim(s) <u>1-9 and 11-34</u> is/are pending in the application.								
-	4a) Of the above claim(s) is/are withdrawn from consideration.							
5)⊠ Claim(s) <u>6,7,19,20 and 30-34</u> is/are allowed.								
6)🛛 (6) Claim(s) <u>1-5,8,9,11-18,21-29</u> is/are rejected.							
-	7) Claim(s) is/are objected to.							
8) 🗌 (8) Claim(s) are subject to restriction and/or election requirement.							
Applicatio	n Papers							
9)□ ⊤	he specification is objected to by the	Examiner.						
• —	•] objected to by the	Examiner.				
10) The drawing(s) filed on is/are: a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).								
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).								
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.								
Priority ur	nder 35 U.S.C. § 119							
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of:								
,	. Certified copies of the priority of	documents have been	received.					
	Certified copies of the priority of			tion No				
	B. ☐ Copies of the certified copies of		* -		Stage			
	application from the International Bureau (PCT Rule 17.2(a)).							
* Se	* See the attached detailed Office action for a list of the certified copies not received.							
Attachment(s)							
1) Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413)								
Notice of Draftsperson's Patent Drawing Review (PTO-948) Paper No(s)/Mail Date								
1 apoi 110(5)/miaii Date								

Art Unit: 3743

DETAILED ACTION

Page 2

Response to Amendment/Arguments

- 1. Applicant's arguments filed 11/10/05 have been fully considered but they are not persuasive.
- 2. In response to Applicant's arguments with respect to the body portion taught by Landis et al, the Examiner contends that the body portion is made of material that is flexible (See Col. 7, lines 44-46, also Col. 9, lines 38-50). Additionally, the device of Landis is fully capable of performing the function set forth in the newly amended claim.

Claim Rejections - 35 USC § 102

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- 4. Claims 1, 3, 5, 8, 9, 11, 13, 16, 17, 21-25, and 27 are rejected under 35 U.S.C. 102(b) as being anticipated by Landis et al (US 5,687,715).
- 5. As to claim 1, Landis et al teaches an apparatus comprising a nose piece 46 comprising a body portion 48; first hollow tube 120 protruding out from the body portion; and a second hollow tube 122 protruding out from the body portion; a first device 44 for attaching the first and second hollow tubes to an individual's head (Please note that Col. 7, lines 28-32 discloses the nasal elements of Fig. 5 can be used with the device configuration of Fig. 1); wherein a first end 50 of the first hollow tube can be inserted into a first nostril of an individual; wherein a first end 52 of the second hollow

Art Unit: 3743

tube can be inserted into a second nostril of the individual; and wherein air can flow through the first hollow tube into the first nostril and through the second hollow tube into the second nostril (See Col. 7, lines 44-54); and wherein the body portion is flexible, wherein the body portion has a primary path and wherein all gas flowing through the body portion to either of the first or second hollow tubes flows through the primary path; wherein the body portion is comprised of a material which is flexible (See Col. 7, lines 44-46, also Col. 9, lines 38-50) and which immediately surrounds the primary path; and wherein the first and second hollow tubes branch out from the primary path to form the first and secondary paths respectively.

Page 3

- 6. As to claim 3, Landis et al teaches an apparatus wherein the body portion is hollow (See Col. 7, lines 44-54).
- 7. **As to claim 4**, Landis et al teaches an apparatus wherein the body portion is flexible (See Col. 7, lines 45-46).
- 8. **As to claim 5**, Landis et al teaches an apparatus wherein the body portion is cylindroid (**See Fig.5**).
- 9. **As to claim 8**, Landis et al teaches an apparatus further comprising a third hollow tube **16**; a fourth hollow tube **18**; and wherein the third and fourth hollow tubes protrude out from the body portion of the nosepiece.
- 10. **As to claim 9**, Landis et al teaches an apparatus wherein the first and second hollow tubes have a first diameter **50,52**; and the third and fourth hollow tubes have a second diameter (**See Fig. 1**); and wherein the first diameter is substantially smaller

Art Unit: 3743

than the second diameter (See Fig. 5 compared to Fig. 1 which shows diameter of 50,52 to be smaller than that of 16,18).

Page 4

- 11. As to claim 11, Landis et al teaches an apparatus further comprising a fifth hollow tube 32 and a sixth hollow tube 34; and wherein the fifth and sixth hollow tubes are connected to the third and fourth hollow tubes, respectively (See Fig. 1); and further comprising a seventh hollow tube 14 and an eight hollow tube 12; wherein the seventh and eighth hollow tubes are connected to the fifth and sixth tubes respectively and the seventh and the eighth hollow tubes are attached to a second device 24,30 for attaching the apparatus to the individual's head; and wherein the first device 44 is adapted to attach the apparatus at or near the upper lip of the individual and the second device is adapted to attach the apparatus at or near the forehead of the individual (See Figs. 1 and 5)
- 12. **As to claim 13**, Landis et al teaches a ventilator circuit (**See Col. 6, lines 2-9**) for supplying air to and through the first and second hollow tubes, and to the first and second nostrils.
- 13. As to claim 16, Landis et al teaches a method comprising the steps of attaching first 120 and second hollow tubes 122 to an individual's head; inserting a first end 50 of the first hollow tube into a first nostril of an individual; and inserting a first end 52 of the second hollow tube into a second nostril of the individual; and wherein the first and second hollow tubes protrude out from a body portion 48 of a nose piece 46 and wherein the first and second hollow tubes are attached to the individual's head through the nose piece, which is attached at or near the upper lip of the individual (See Fig. 1);

Art Unit: 3743

and wherein the body portion of the nose piece is flexible (See Col. 7, lines 45-46);); and wherein the body portion is flexible, wherein the body portion has a primary path and wherein all gas flowing through the body portion to either of the first or second hollow tubes flows through the primary path; wherein the body portion is comprised of a material which is flexible (See Col. 7, lines 44-46, also Col. 9, lines 38-50) and which immediately surrounds the primary path; and wherein the first and second hollow tubes branch out from the primary path to form the first and secondary paths respectively.

Page 5

- 14. **As to claim 17**, Landis et al teaches a method further comprising connecting a third hollow tube **16** to the first hollow tube; and connecting a fourth hollow tube **18** to the second hollow tube; attaching the third hollow tube at or near the forehead of an individual; and attaching the fourth hollow tube at or near the forehead of an individual (**See Fig. 1**).
- 15. **As to claim 21**, Landis et al teaches a method further comprising attaching third **16** and fourth hollow tubes **16** to the body portion of the nose piece; and wherein the third and the fourth hollow tubes protrude out from the body portion of the nose piece.
- 16. As to claim 22, Landis et al teaches a method wherein the first and second hollow tubes have a first diameter 50,52; and the third and fourth hollow tubes have a second diameter (that of 16 and 18); and wherein the first diameter is substantially smaller than the second diameter (See Fig. 5 compared to Fig. 1 which shows diameter of 50,52 to be smaller than that of 16,18).

Art Unit: 3743

17. **As to claim 23**, Landis et al teaches a method further comprising attaching a fifth hollow tube **32** to the third hollow tube; and attaching a sixth hollow tube **34** to the fourth hollow tube.

- 18. **As to claim 24**, Landis et al teaches a method further comprising supplying air (**via air source depicted in Fig. 1**) to the fifth hollow tube and thereby to the third, and first hollow tubes and to the first nostril; and supplying air to the sixth hollow tube and thereby to the fourth, and second hollow tubes and to the second nostril.
- 19. **As to claim 25**, Landis et al teaches a method further comprising attaching a seventh hollow tube **14** to the fifth hollow tube; attaching an eighth hollow tube **12** to the sixth hollow tube; and attaching the seventh hollow tube and eighth hollow tubes at or near the individual's forehead (**See Fig. 1**).
- 20. As to claim 27, Landis et al teaches a method further comprising supplying air to and through the first and second hollow tubes, and to the first and second nostrils (See Col. 7, lines 44-54).

Claim Rejections - 35 USC § 103

- 21. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Art Unit: 3743

22. The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

- 1. Determining the scope and contents of the prior art.
- 2. Ascertaining the differences between the prior art and the claims at issue.
- 3. Resolving the level of ordinary skill in the pertinent art.
- 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.
- 23. Claims 2, 4, 12, 14, 15, 18, 26, 28 and 29 are rejected under 35 U.S.C. 103(a) as being unpatentable over Landis et al.
- 24. **As to claims 2, 4, and 18**, Landis et al teaches essentially all the limitations except for the specific material of the body portion. However, it would have been obvious to one of ordinary skill in the art at the time the invention was made to make the body portion of the claimed material, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin, 125 USPQ 416.*
- 25. As to claims 12, 14, 26, and 28, Landis et al teaches essentially all of the limitations except for the apparatus further comprising ninth and tenth hollow tubes. However, it would have been obvious to one of ordinary skill in the art at the time of the invention to provide ninth and tenth hollow tubes, since it has been held that mere duplication of the essential working parts of a device involves only routine skill in the art. St. Regis Paper Co. I v. Bemis Co., 193 USPQ 8.

Art Unit: 3743

26. As to claims 15 and 29, Landis et al teach essentially all of the limitations except for the use of an adhesive. Landis et al rather teaches the use of a headband strap. Applicant has not disclosed that the adhesive solves a stated problem or provides unexpected results. One of ordinary skill in the art would have expected the device of Landis with the headband strap and Applicant's invention with the use of an adhesive to work equally as well since both function to provide a means of securement of the device to the user's head.

Allowable Subject Matter

- 27. Claims 6, 7, 19, 20, 30, 31, 32, 33 and 34 are allowed over the prior art of record.
- 28. The following is a statement of reasons for the indication of allowable subject matter: As to claim 33, the prior art of record does not teach nor render obvious the overall claimed combination of an apparatus for use in supplying air to an individual comprising a means for attaching the first device to an individual's forehead; a means for attaching the second device to an area at or near the individual's lip; wherein the means for attaching the first device does not circle a head of the individual in order to attach the first device; and wherein the means for attaching the second device does not circle the head of the individual in order to attach the second device.

Conclusion

29. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

Art Unit: 3743

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Mital B. Patel whose telephone number is 571-272-4802. The examiner can normally be reached on Monday-Friday (11:00-7:30).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Henry Bennett can be reached on 571-272-4791. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3743

Page 10

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Mital B. Patel Primary Examiner Art Unit 3743